

X-RAY OUTFITS FOR HOSPITAL OF SEVENTY-FIVE BEDS OR LESS

EQUIPMENT

United States Army bedside unit.....	1
Coolidge tube, radiator type.....	2
Rotary converter that will start on 110 volts without requiring a starting resistance. Must operate unit as on 110-volt alternating current.	
Fluoroscope, 11 by 14, special folding, with demountable stand and detachable plate-holding device as furnished by Waite & Bartlett.....	1
Fluoroscope, hand, with 1-piece extra lead glass.....	1
Gloves, gauntlet, lead rubber.....	2
Auto transformer, 1½ kilowatts, 110 to 220.....	1
10-ampere screw fuse plugs.....	4
Extra set of brushes for rotary converter.....	1
10-ohm resistance for 220-volt rotary.....	1
Extra lead glass shield.....	1
Extra reel.....	1
Goggles, automobile type, fitted with red and green superimposed clear polished celluloid, Arlington Pyralin in colors 24 and 181, 0.01 inch thick, finish H. H.....	1 pairs
Foot switch (W. & B. dental).....	1
Apron, protective.....	1
Box, illumination, 8 by 10 inches.....	1

Chemicals:

Chrome alum.....	pounds..	1
Hydroquinone.....	do.....	2
Metol.....	ounces..	4
Potassium bromide.....	pounds..	1
Sodium carbonate, dry.....	do.....	2
Sodium hyposulphite.....	do.....	10
Sodium sulphite, dry.....	do.....	2

Films:

Dental, negative, 1¼ by 1½ inches.....	gross..	1
X-ray—		
5 by 7 inches, Eastman Kodak Co.....	dozen..	10
8 by 10 inches, Eastman Kodak Co.....	do.....	10
Gloves, opaque, protective, gauntlet.....	pair..	1
Pitcher, enamel, 2-quart.....		1
Safe lights, Eastman Brownie.....		2
Screens, intensifying, aluminum cassettes:		
5 by 7 inches.....		4
8 by 10 inches.....		2
Trays, enamel ware:		
14 by 17 inches.....		2
8 by 10 inches.....		2